

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Essex Fells P.D., Essex Fells, NJDate Received: 06/07/2017Date given to service: 06/07/2017Carrier: ☒ FedEx

UPS

USPS

Shipping Method: ☒ GRD

3DAY

2DAY

NDA-PR

NDA-STD

Product: 6510

6810

☒ 7110

8610

Serial Number: AR XD-U009

7510

DT5000

Printer Serial#: AR

7410 Upper-half

9510

Sim Serial#: DD

7410 Whole

Probe Serial#: DD

Warranty Expires: _____

Description:

A

B

Plus

Demo

Screener

Trade In

Accessories

110V A/C Adapter

Regulator

Printer Ribbon

Printer Paper

Mouthpieces

9510 Stylus

9510 Top Cover

Carrying Case

Dry Gas

Other (please specify) _____

Repair Information.

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration		
4414166	Labor	1	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell, geared engine, and spacer plates.

Fuel Cell = ARKE-0192

CAL W/QC AND OPS CHECK

Service Technician: BCDate: 08/02/17

4
31
A
Form 27
revision 10/01/01
Public Form
return and res.