

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Palisades Interstate Parkway Police, Alpine, NJ

Date Received: 06/14/2016 Date given to service: 06/14/2016

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR XD-0010
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: January 2014 (77931)

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	1	
6808455	Fuel Cell	1	

Repair Notes: Replaced defective fuel cell, C156 and C167.
C156 and C167 were changed due to modem issues.

Fuel Cell = ARJD-1963
 CAL W/QC AND OPS CHECK

Service Technician BC Date: 07/07/2016