

**Return & Repair Form:**

**Customer Information:**

Company Name: Irvington Police B: 78747 S: \_\_\_\_\_  
 Date Received: 5-24-10 Date Given to Service: \_\_\_\_\_  
 Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-~~7110~~-6510-6810-7510  
 9510- Drug Tester 5000 Serial #: AR XE-0075  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR \_\_\_\_\_  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P \_\_\_\_\_  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date Dec. 2013

**Repair Information:**

Test #

Reason for Return:			
<u>FC</u>			
Part Number	Description	Qty	Total Cost
<u>6808955</u>	<u>FC</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes:

NEW FC # ARAN-1017  
Cal w/Op & Ops Check

Service Technician: [Signature]

Date: 05-24-2010