

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Bogota P.D., Bogota NJ

Date Received: 06/06/13

Date given to service: 06/06/13

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR ZL-0131

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2014

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter
☐ Mouthpieces
☐ Dry Gas

☐ Regulator
☐ 9510 Stylus
☐ Other (please specify) _____

☐ Printer Ribbon
☐ 9510 Top Cover

☐ Printer Paper
☐ Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = AREB-1808

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 06/12/13