

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Rochelle Park P.D., Rochelle Park, NJ

Date Received: 09/08/14

Date given to service: 09/08/14

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR ZL-0135

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2014

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	8	
8315075	Alcotest 7110 Motherboard	1	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	
6808374	Geared Engine	1	

Repair Notes: Replaced defective Alcotest 7110 Motherboard due to BLOWING NOT ALLOWED.

Replaced defective geared engine due to not functioning properly and getting stuck.

Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFH-1964

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 10/14/14