

# Return & Repair Form:

## Customer Information:

Company Name: Bard of Wood-Ridge B: 19596 S: \_\_\_\_\_

Date Received: 7-7-10 Date Given to Service: \_\_\_\_\_

Carrier: FedEx UPS DHL - USPS Method: GRD NDA - 2DY - 3DY - Other \_\_\_\_\_

## Product Information:

Product: 7410-~~7110~~6510-6810-7510  
9510- Drug Tester 5000

Serial #: AR 2L-0139

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_

Probe Ser # DD \_\_\_\_\_ P -

## ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# \_\_\_\_\_

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # \_\_\_\_\_

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) \_\_\_\_\_

Warranty Exp. Date JAN-2014

## Repair Information:

Test #

Reason for Return: FC

| Part Number     | Description | Qty       | Total Cost   |
|-----------------|-------------|-----------|--------------|
| <u>6808455</u>  | <u>FC</u>   | <u>1</u>  | <u>N/C w</u> |
| <u>MPCAL 71</u> | <u>Cal</u>  | <u>1</u>  | <u>N/C w</u> |
| <u>MP Laba</u>  | <u>Labo</u> | <u>.5</u> | <u>N/C w</u> |

Repair Notes:

Now FC # ARBD-0118

Cal w/ QC / ops check

Service Technician: [Signature]

Date: 7-8-2010