

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Boro. of Wood-Ridge P.D., Wood-Ridge, NJ

Date Received: 06/24/14

Date given to service: 06/24/14

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR ZL-0139

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2014 (79596)

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	4	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	
6809500	Cuvette	1	

Repair Notes: Replaced cuvette, fuel cell, and spacer plates.

Changed cuvette for Error 23 and BLOWING NOT ALLOWED issue.

Fuel Cell = ARFD-1940

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 08/11/14