

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Wood Ridge P.D., Wood Ridge, NJ

Date Received: 01/11/2018

Date given to service: 01/11/2018

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610 Serial Number: AR ZL-0139
☐ 7510 ☐ DT5000 Printer Serial#: AR
☐ 7410 Upper-half ☐ 9510 Sim Serial#: DD
☐ 7410 Whole Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	1	
6808455	Fuel Cell	1	
6808374	Geared Engine	1	

Repair Notes: Replaced defective fuel cell and geared engine.

Fuel Cell = ARKF-1443

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 01/24/2018