

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Saddle Brook P.D., Saddle Brook NJ

Date Received: 03/01/12 Date given to service: 03/01/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole
Serial Number: AR ZL-0143
Printer Serial#: AR
Sim Serial#: DD
Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>n/c w</u>
<u>mp cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>n/c w</u>
<u>mp Labor</u>	<u>Labor</u>	<u>5</u>	<u>n/c w</u>

Repair Notes: Replaced busted Fuel Cell # AR25-0149
With Fuel Cell # ARCK-0883.

Cal w/RC; ops check

Service Technician: [Signature] Date: 3/17/12