

# RETURN AND REPAIR FORM

# Dräger

## Customer Information

B:

S:

Company Name: Park Ridge P.D., Park Ridge, NJ

Date Received: 04/08/15

Date given to service: 04/08/15

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR ZL-0145

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
8315075	Alcotest 7110 Motherboard	1	NC/W

Repair Notes: Replaced defective motherboard, fuel cell, and spacer plates.

Unit was giving "BLOWING NOT ALLOW" Error after fuel cell was changed and by changing  
motherboard, error was resolved.

Fuel Cell = ARFN-0044

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 05/06/15